

25/03/2026

Notification of change**Product Name:**

Cat.no. E2007 ePure STD DNA Extraction kit

Content of the change:

The expiry date is extended to "24 months" for catalog numbers:
E2007, lot: 231007D

Dear valued customers,

ePure kits now have an extended shelf life, a stability study has been performed and the conclusion is that the expiration date can be extended to 24 months.

The kit from lot 231007D (E2007) was produced after the evaluation lot for stability study, so this lot has the same process, material and quality system, therefore the expiry date for this kit can be extended to 24 months from the date of production.

As confirmed by the internal stability study and under the conditions of proper storage at the temperature specified in the relevant IFU, the stability of the lot below is as follows:


Cat. No.	Product Name	lot	Original expiry date	Extended expiry date
E2007	ePure STD DNA Extraction kit,	231007D	30.09.2026	31.12.2026

Please forward this information to all relevant contacts that may need to know about this change.

We apologize for any inconvenience and thank you for your understanding.

Please contact the email address below for more information:

ecolidx@ecolidx.com

 **ECOLI**
Ecoli Dx, s.r.o.
Purkyňova 74/2
110 00 Praha 1 - Nové Město
IČ: 107 07 409 / DIČ: CZ10707409
Tel.: +420 325 209 912

RNDr. Tereza Kassay

Head of the division
Ecoli Dx s.r.o.
Purkyňova 74/2
110 00 Praha 1