

Instruction Manual

KEY TO SYMBOLS USED

	Catalogue number		Caution
	Batch code		Contains sufficient for <n> tests
	Research Use Only		Use-by Date
	Version		Consult instructions for use
	Temperature limit		Keep away from sunlight
	Manufacturer	NCA	Negative control of amplification
	Date of manufacture	C-	Negative control of extraction
IC	Internal control	C+	Positive control of amplification

1. INTENDED USE

AmpliSens[®] MTC-FRT PCR kit is an *in vitro* nucleic acid amplification test for qualitative detection of *Mycobacterium tuberculosis* (MTB) DNA – *Mycobacterium tuberculosis* complex (MTC), including *M.tuberculosis*, *M.bovis*, *M.africanum*, *M.microti*, *M.canetti*, *M.pinnipedii* – in biological material, cultures of microorganisms and environmental objects using real-time hybridization-fluorescence detection of amplified products.

NOTE: For research use only. Not for diagnostic procedures.

2. PRINCIPLE OF PCR DETECTION

Mycobacteria tuberculosis detection by the polymerase chain reaction (PCR) is based on the amplification of a pathogen genome specific region using specific *Mycobacteria tuberculosis* primers. In the real-time PCR, the amplified product is detected with the use of fluorescent dyes. These dyes are linked to oligonucleotide probes which bind specifically to the amplified product during thermocycling. The real-time PCR monitoring of fluorescence intensities during the real-time PCR allows the detection of accumulating product without re-opening the reaction tubes after the PCR run.

AmpliSens[®] MTC-FRT PCR kit is a qualitative test that contains the Internal Control (Internal Control STI-87). It must be used in the extraction procedure in order to control the extraction process of each individual sample and to identify possible reaction inhibition.

AmpliSens[®] MTC-FRT PCR kit uses "hot-start", which greatly reduces the frequency of nonspecifically primed reactions. "Hot-start" is guaranteed by separation of nucleotides and Taq-polymerase by using a wax layer or a chemically modified polymerase (TaqF). Wax melts and reaction components mix only at 95 °C. Chemically modified polymerase (TaqF) is activated by heating at 95 °C for 15 min.

AmpliSens[®] MTC-FRT PCR kit includes enzyme uracil-DNA glycosylase (UDG) to reduce the risk of contamination.

For optimization of *Mycobacteria tuberculosis* research report, an integrated procedure of DNA extraction for quantitative detection, identification to species, and determination of resistance to antitubercular therapy can be carried out.

The results of amplification are registered in the following fluorescence channels:

Table 1

Channel for fluorophore	FAM	JOE
DNA-target	DNA <i>M tuberculosis</i> complex	DNA IC
Target gene	IS 6110	Artificially synthesized sequence

3. CONTENT

AmpliSens[®] MTC-FRT PCR kit is produced in 2 forms:

variant FRT,  R-B57(RG,iQ,SC,Dt)-CE;

variant FRT in bulk¹,  R-B57(RG,iQ,SC,Dt)-CE-B.

Variant FRT includes:

Reagent	Description	Volume, ml	Quantity
PCR-mix-1-FRT MTC	clear liquid from colorless to light lilac colour	0.28	2 tubes
PCR-buffer-Flu	colorless clear liquid	0.28	1 tube
Polymerase (TaqF)	colorless clear liquid	0.03	1 tube
Enzyme UDG	colorless clear liquid	0.03	1 tube
Positive Control DNA MTC / STI (C+MTCSTI)	colorless clear liquid	0.1	1 tube
TE-buffer	colorless clear liquid	0.5	1 tube
Negative Control (C-)*	colorless clear liquid	1.6	1 tube
Internal Control STI-87 (IC)**	colorless clear liquid	1.0	1 tube
RNA-buffer***	colorless clear liquid	1.2	1 tube

* must be used in the extraction procedure as Negative Control of Extraction.

** add 10 µl of Internal Control STI-87 (IC) during the DNA extraction procedure directly to the sample/lysis mixture (see DNA-sorb-C or DNA-sorb-B or RIBO-prep protocols).

*** used for elution during DNA extraction (for RIBO-prep).

Variant FRT is intended for 55 reactions (including controls).

4. ADDITIONAL REQUIREMENTS

- Reagent for pretreatment of viscous fluids (sputum, aspirates).
- Homogenizer is recommended to use for tissue material homogenization.
- Sterile stainless steel balls with 5 mm and 7 mm diameter for tissue material homogenization.
- Sterile porcelain or glass beads with 3-5 mm diameter for sputum homogenization and 3 mm diameter for tissue material homogenization.
- DNA extraction kit.
- Disposable powder-free gloves and laboratory coat.
- Pipettes (adjustable).
- Sterile pipette tips with aerosol barriers (up to 200 µl).
- Tube racks.
- Vortex mixer.
- Desktop centrifuge with rotor for 2 ml reaction tubes.
- PCR box.
- Real-time instruments (for example, Rotor-Gene 3000/6000 (Corbett Research, Australia); iCycler iQ or iCycler iQ5 (Bio-Rad, USA); Mx 3000P, Mx3005 (Stratagene, USA); SmartCycler II (provided with Mini-Spin centrifuge) (Cepheid, USA) or equivalent).
- Disposable polypropylene PCR tubes:
 - a) 0.2-ml PCR tubes (flat caps, nonstriped) for 36-well rotor if a rotor-type instrument is used;
 - b) 0.2-ml PCR tubes (domed caps) if a plate-type instrument is used.
- Refrigerator for 2–8 °C.
- Deep-freezer at the temperature from minus 24 to minus 16 °C.
- Reservoir for used tips.

¹ In bulk form contains unlabeled tubes. Tubes with identical reagent are packed in one bag with label.

5. GENERAL PRECAUTIONS

The user should always pay attention to the following:

- Use sterile pipette tips with aerosol filters and use a new tip for every procedure.
- Store all extracted positive material (specimens, controls and amplicons) away from all other reagents and add it to the reaction mix in a distinctly separated facility.
- Thaw all components thoroughly at room temperature before starting an assay.
- When thawed, mix the components and centrifuge briefly.
- Use disposable protective gloves and laboratory cloths, and protect eyes while samples and reagents handling. Thoroughly wash hands afterwards.
- Do not eat, drink, smoke, apply cosmetics, or handle contact lenses in laboratory work areas.
- Do not use a kit after its expiration date.
- Dispose of all specimens and unused reagents in accordance with local regulations.
- Samples should be considered potentially infectious and handled in biological cabinet in compliance with appropriate biosafety practices.
- Clean and disinfect all samples or reagents spills using a disinfectant, such as 0.5 % sodium hypochlorite or another suitable disinfectant.
- Avoid inhalation of vapors, samples and reagents contact with the skin, eyes, and mucous membranes. Harmful if swallowed. If these solutions come into contact, rinse the injured area immediately with water and seek medical advice if necessary.
- Safety Data Sheets (SDS) are available on request.
- Use of this product should be limited to personnel trained in DNA amplification techniques.
- Workflow in the laboratory must be one-directional, beginning in the Extraction Area and moving to the Amplification and Detection Area. Do not return samples, equipment and reagents in the area where the previous step was performed.



Some components of this kit contain sodium azide as a preservative. Do not use metal tubing for reagent transfer.

6. SAMPLING AND HANDLING

Sampling

- 6.1 *Bronchoalveolar lavage (BAL)* and *bronchoalveolar lavage fluid (BALF)*, *liquor*, are collected to disposable hermetically screwed polypropylene vessels (for preventing adhesion of the cells on their internal surface) with a volume no less than 5 ml.
- 6.2 *Sputum* and *urine (medium portion)* is collected to disposable graduated screwed vessels with a wide neck with a volume no less than 50 ml.
- 6.3 *Fasting morning whole blood* and *pleural fluid* is collected to tubes (for example, Vacuette®) with EDTA spraying or its solute. Close the tube and turn it upside down and back several times.
- 6.4 *Menstrual blood* is collected to dry disposable test tubes using a Kafka cap.
- 6.5 *Synovial fluid* is collected to dry disposable test tubes.
- 6.6 *Prostate gland secretion* is collected to sterile disposable 1.5-ml tubes after massage of the prostate gland. If, after massage of prostate gland, it is impossible to get the secretion, use the first portion of urine, which contains the prostate gland secretion.
- 6.7 *Tissue (biopsy, surgical) material* is collected to tubes (for example, Vacuette®) with EDTA spraying or to disposable tubes with 0.2 ml of sterile saline or PBS.
- 6.8 *Paraffin units* are cut by using microtome or cut out a fragment of tissue by disposable scalpel. Then remove paraffin by using xylene, remove xylene by series of ablation with decrease of ethanol concentration (similarly to standard histology conducting).
- 6.9 *Cultures of microorganisms* grown on selective solid nutrient media for *Mycobacteria tuberculosis* are collected to glass tubes as working with turbidity standard by resuspending in saline. *Cultures of microorganisms* grown on selective liquid nutrition media are used in original vial.
- 6.10 *Washing fluids from environmental objects* are collected with a tent with a wad wetted in saline. The square of washing from flat surface is 5-10 cm². The working part of the tent is to be transferred to the 1.5-ml tube with 0.5 ml of sterile saline. The top of the tent is to be broken and removed.

The samples (except for urine) are to be stored at 2–8 °C for 3 days, at –16 °C for 1 year. For archiving (more than 1 year), store the samples at –68 °C. Urine can be stored at 2–8 °C for no longer than 6 hours. Freeze urine for a long storage. Double freezing-thawing of the biological material is allowed.

NOTE: Do not freeze blood.

Transport the samples in thermocontainer for no more than 3 days.

Pretreatment

- 6.11 Mix *BAL* or *BALF* by turning upside down and back. Transfer 1 ml of the sample to a 1.5-ml Eppendorf tube using a pipette with a tip with aerosol barrier, mark it, and centrifuge at 10000 g for 10 min. Carefully remove the supernatant using a tip with aerosol barrier and leaving about 100 µl of the sample.
- 6.12 Add "Mucolysin" to the vessel with *sputum* (5 : 1, v/v) and then add and 3-5 sterile porcelain or glass beads to this mixture. Stir the vessel periodically for 20–30 min. Transfer 100 µl of the sample to 1.5-ml tube Eppendorf using a pipette with a tip with aerosol barrier and mark it.
- 6.13 Mix *urine* by turning the vessel upside down and back. Using a pipette with a tip with aerosol barrier, transfer 5–10 ml of the sample to a screwed tube, mark it, and centrifuge at 10000 g for 10 min (or at 3000 g for 20 min). Carefully remove the supernatant using a tip with aerosol barrier and leaving about 100 µl of the sample (if the pellet is visible, remove the supernatant leaving just a pellet).
- 6.14 Add "Mucolysin" to the vessel with *synovial fluid* (1:1, v/v). Stir the vessel periodically for 20-30 min.
- 6.15 Transfer *tissue material* to a disposable Petri dish. Mince the fragment (10 mm³ or 10 µl) with a disposable scalpel. If a 12-well Multispin MSC-6000 vortex/centrifuge (BioSan, Latvia) is used, transfer fragments of tissue to 2-ml disposable screwed polypropylene conical tubes with loops and add 2 or 3 sterile glass beads. If a TissueLyser LT homogenizer (QIAGEN, Germany) is used, transfer tissue fragments to 2-ml disposable screwed tubes and add 1 or 2 sterile metal beads. If porcelain mortars and pestles are used, transfer tissue fragments to a mortar and add an equal volume of PBS or sterile saline. Homogenize the sample.
- 6.16 Resuspend *cultures of microorganisms grown on solid nutrient medium (SNM)* in a sterile saline or PBS using turbidity standard No. 5 (5x10⁸ microbial bodies per 1 ml (m.b./ml)) or McFarland No. 0.5, 1 or 2. Use 5 µl of this suspension. Take a 1-ml aliquot of *cultures of microorganisms grown on liquid nutrient medium (LNM)* and centrifuge it at 1000 g for 10 min. Discard the supernatant.
- 6.17 Use 100-µl aliquots of *washing fluids from environmental objects*.

The samples volume for treatment and DNA extraction

Material	Aliquot volume for treatment	Aliquot volume for DNA extraction
Sputum	All sample	0.1 ml
BAL or BALF	1 ml	0.1 ml
Urine	5–10 ml	0.1 ml
Liquor	1 ml	0.1 ml
Synovial fluid	1 ml	0.1 ml
Prostate gland secretion	1 ml	0.1 ml
SNM	1.5-6 x 10 ⁹ m.b./ml	0.05 ml
LNM	1 ml	0.1 ml
Blood		0.1 ml
Menstrual blood		0.1 ml
Tissue		10–25 µl
Washing fluids from environmental objects		0.1 ml

NOTE: It is necessary to prevent the repeated sample extraction and reserve the sample aliquot in accordance with storage regulations.

7. WORKING CONDITIONS

AmpliSens® MTC-FRT PCR kit should be used at 18–25 °C.

8. PROTOCOL

8.1. DNA extraction

It is recommended to use the following nucleic acid extraction kits:

- **DNA-sorb-B** (for biological material, cultures of microorganisms, and environmental objects).
- **RIBO-prep** (for biological material, cultures of microorganisms, and environmental objects).
- **DNA-sorb-C** (for human tissues).

DNA extraction of each test sample is carried out in the presence of **Internal Control STI-87 (IC)**.

In the extraction procedure it is necessary to carry out the control reaction as follows:

- C–** – Add **100 µl of Negative Control (C–)** to the tube labelled C– (Negative Control of Extraction).

NOTE: **If using DNA-sorb-B**, extract the DNA according to the manufacturer's protocol taking into account next additions and improvements:

- In case of DNA extraction from the urine add the **Lysis Solution** into the tubes with material pellet, resuspend it using individual tip for each sample and transfer into the 1.5-ml Eppendorf tubes.
- Carry out DNA elution in **100 µl of TE-buffer for DNA elution**

If using RIBO-prep kit, extract the DNA according to the manufacturer's protocol taking into account next additions and improvements:

- In case of DNA extraction from the urine add the **Lysis Solution** into the tubes with material pellet, resuspend it using individual tip for each sample and transfer into the 1.5-ml Eppendorf tubes.
- Carry out DNA elution in **100 µl of RNA-buffer**

NOTE:

If using DNA-sorb-C kit, extract the DNA according to the manufacturer's protocol taking into account next additions and improvements:

- In case of using a Multispin MSC-6000 vortex/centrifuge (BioSan, Latvia) add **400 µl of Lysis Reagent Buffer**, **17 µl of Lysis Reagent** and **10 µl of Internal Control STI-87 (IC)** to each tube with tissue material and beads. Homogenization mode:

RPM	Vortex	Cycle
1000	hard	140
0,01	20	stop

Insert the tubes into the thermostat and incubate it at the temperature 65 °C for 10 min.

NOTE:

- In case of using an automatic homogenizer TissueLyser LT (QIAGEN, Germany) add **400 µl of Lysis Reagent Buffer**, **17 µl of Lysis Reagent** and **10 µl of Internal Control STI-87 (IC)** to each tube with tissue material and beads. Homogenization mode: frequency – 50 Hz, time of homogenization – 2 min.
 - Insert the tubes into the thermostat and incubate it at the temperature 65 °C for 10 min.
 - In case of using porcelain mortars add **400 µl of Lysis Reagent Buffer**, **17 µl of Lysis Reagent**, **10 µl of Internal Control STI-87 (IC)** and **20 µl of homogenized tissue sample** into each tube.
- Incubate the samples at the temperature 65 °C for 1 hour, shaking it occasionally (at least 5 times).
- Carry out DNA elution in **100 µl of TE-buffer for DNA elution**.

8.2. Preparing PCR

8.2.1 Preparing tubes for PCR

The total reaction volume is **25 µl**, the volume of DNA sample is **10 µl**.

Use disposable filter tips for adding reagents, DNA and control samples into tubes.

Before starting work, thaw and thoroughly vortex all reagents of the kit. Make sure that there are no drops on the caps of the tubes. All components of the reaction mixture should be mixed immediately before use.

NOTE:

1. Take the required number of tubes for amplification of the DNA obtained from test and control samples.
2. For N reactions, add to a new tube:
 - 10*(N+1) µl of PCR-mix-1-FRT MTC,**
 - 5*(N+1) µl of PCR-buffer-Flu,**
 - 0.5*(N+1) µl of polymerase (TaqF),**
 - 0.5*(N+1) µl of enzyme UDG.**
 Vortex the tube, then centrifuge it briefly
3. Transfer **15 µl** of the reaction mixture to each tube.

4. Add 10 µl of DNA samples obtained at the DNA extraction stage from test or control samples.
NOTE Avoid transferring sorbent beads together with the DNA.
5. Carry out the control amplification reactions:
NCA – Add 10 µl of TE-buffer to the tube labeled NCA (Negative Control of Amplification).
C+ – Add 10 µl of Positive Control DNA MTC/ STI to the tube labeled C+ (Positive Control of Amplification).
C- – Add 10 µl of the sample extracted from the Negative Control (C-) reagent to the tube labeled C- (Negative control of Extraction).

NOTE: For carrying out decontamination of the reaction mixture incubate prepared tubes at room temperature for 10–30 min.

8.2.2 Amplification

1. Create a temperature profile on your instrument as follows:

Table 3

Step	Temperature, °C	Time	Fluorescence detection	Cycles
1	95	15 min	–	1
2	95	15 s	–	5
	65	30 s	–	
	72	15 s	–	
3	95	15 s	–	40
	65	30 s	FAM/Green, JOE/Yellow, ROX/Orange, Cy5/Red	
	72	15 s	–	

Table 4

Step	Temperature, °C	Time	Fluorescence detection	Cycles
1	95	15 min	–	1
2	95	15 s	–	5
	65	30 s	–	
	72	15 s	–	
3	95	15 s	–	40
	65	30 s	FAM, JOE/HEX, ROX, Cy5	
	72	15 s	–	

Table 5

Step	Temperature, °C	Time	Fluorescence detection	Cycles
1	95	15 min	–	1
2	95	20 s	–	45
	65	50 s	Optics ON	
	72	20 s	–	

- Adjust the fluorescence channel sensitivity according to the *Important Product Information Bulletin* and Guidelines [1].
 - Insert tubes into the reaction module of the device.
- NOTE** It is recommended to sediment drops from walls of tubes by short centrifugation (1–3 s) before placing them into the instrument.
- Run the amplification program with fluorescence detection.
 - Analyze results after the amplification program is completed.

9. DATA ANALYSIS

Analysis of results is performed by the software of the real-time PCR instrument used by measuring fluorescence signal accumulation in two channels:

- The signal of the IC DNA amplification product is detected in the channel for the JOE fluorophore,
- The signal of the *Mycobacterium tuberculosis complex* DNA amplification product is detected in the channel for the FAM fluorophore.

The results are interpreted by the crossing (or not-crossing) of the fluorescence curve with the threshold line set at the specific level that corresponds to the presence (or absence) of a Ct value of the DNA sample in the corresponding column of the results grid. Principle of interpretation is specified in the Table 6.

Table 6

Ct value in the channel for fluorophore		Result validity	Interpretation
FAM	JOE		
< boundary value	< boundary value or Absent	Valid	<i>M.tuberculosis complex is detected</i>
Absent	< boundary value	Valid	<i>M.tuberculosis complex is not detected</i>
Absent or > boundary value	Absent or > boundary value	Invalid	Invalid result (repeat material sampling and assay)
> boundary value	< boundary value	Invalid	Equivocal result (repeat material sampling and assay)

NOTE Boundary Ct values are specified in the *Important Product Information Bulletin* enclosed to the PCR kit. See also Guidelines [1]

- If the result is positive in the channel for FAM fluorophore (the Ct value does not exceed the boundary Ct value, specified in the *Important Product Information Bulletin*) and the result is positive (the Ct value does not exceed the boundary Ct value, specified in the *Important Product Information Bulletin*) or negative (Ct value is absent) in the channel for JOE fluorophore - the result is **valid**, *Mycobacterium tuberculosis* DNA is **detected**.
- If the result is negative (Ct value is absent) in the channel for FAM fluorophore and the result is positive (the Ct value does not exceed the boundary Ct value, specified in the *Important Product Information Bulletin*) in the channel for JOE fluorophore - the result is **valid**, *Mycobacterium tuberculosis* DNA is **not detected**.
- If the result is negative (Ct value is absent) or the Ct value exceeds the boundary Ct value, specified in the *Important Product Information Bulletin* in the channels for JOE and FAM fluorophores, the result is **invalid**. It is necessary to repeat amplification. If the result is the same, repeat the assay beginning from the DNA extraction. If the result is the same again, it is considered to be **invalid**. In this case, it is recommended to repeat material sampling and assay.

² For example, Rotor-Gene 3000/ 6000 (Corbett Research, Australia), Rotor-Gene Q (QIAGEN, Germany).

³ For example, iCycler iQ, iQ5 (Bio-Rad, USA).

- If the Ct value exceeds the boundary Ct value, specified in the *Important Product Information Bulletin* in the channel for FAM fluorophore and the result is positive (the Ct value does not exceed the boundary Ct value, specified in the *Important Product Information Bulletin*) in the channel for JOE fluorophore, the result is **invalid**. It is necessary to repeat amplification. If the result is the same, repeat the assay beginning from the DNA extraction. If the result is the same again, it is considered to be **equivocal**. In this case, it is recommended to repeat material sampling and assay
- The result of the analysis is considered reliable only if the results obtained for Positive and Negative Controls of amplification as well as for the Negative Control of extraction are correct (see Table 7).**

Table 7

Control	Stage for control	Ct value in the channel for fluorophore	
		FAM	JOE
C-	DNA extraction	Absent	< boundary value
C+	PCR	< boundary value	< boundary value
NCA	PCR	Absent	Absent

10. TROUBLESHOOTING

The results of the analysis are not taken into account in the following cases:

- If the positive signal is absent for the Positive Controls of amplification in the channels for FAM or JOE fluorophores, it may suggest that the wrong amplification program was chosen, or other mistakes were made in the amplification stage. The PCR amplification should be repeated for all negative samples. If the same result is obtained again, the PCR analysis (beginning with the DNA extraction stage) should be repeated for such samples.
- If the positive signal is detected for the Negative Control of Amplification (NCA) and/or Negative Control of Extraction (C-) in the channel for FAM fluorophore, it indicates contamination of reagents or test samples. The PCR analysis (beginning with the DNA extraction stage) should be repeated for all positive samples.
- If the positive signal is absent for the Negative Control of Extraction (C-) in the channel for FAM fluorophore, it indicates the mistakes made in the DNA extraction stage. In this case the PCR analysis (beginning with the DNA extraction stage) should be repeated for the samples.

11. TRANSPORTATION

AmpliSens® MTC-FRT PCR kit should be transported at 2–8 °C for no longer than 5 days.

12. STABILITY AND STORAGE

All components of the **AmpliSens® MTC-FRT** PCR kit are to be stored at 2–8 °C when not in use (except for polymerase (TaqF), enzyme UDG, and PCR-mix-1-FRT MTC). All components of the **AmpliSens® MTC-FRT** PCR kit are stable until the expiration date on the label. The shelf life of reagents before and after the first use is the same, unless otherwise stated.

NOTE: Polymerase (TaqF), enzyme UDG, and PCR-mix-1-FRT MTC are to be stored at the temperature from minus 24 to minus 16 °C.

NOTE: PCR-mix-1-FRT MTC is to be kept away from light.

13. SPECIFICATIONS

13.1. Analytical sensitivity

DNA extraction kit	Material	Sensitivity, mb/ml
		<i>M.tuberculosis</i> (H37 Ra strain)
RIBO-prep	PBS, sputum, BAL	5x10 ²
	Urine	1x10 ³
	Washing fluids from environmental objects ⁴	2.5x10 ² copies/ml
DNA-sorb-B	PBS, sputum	5x10 ²
	BAL, urine	1x10 ³
	Washing fluids from environmental objects	2.5x10 ² copies/ml
DNA-sorb-C	10 % homogenate of different tissues (lungs, lymph nodes, kidney, liver, brain, spleen)	1x10 ²

13.2. Analytical specificity

The analytical specificity of **AmpliSens® MTC-FRT** PCR kit is ensured by selection of specific primers and probes as well as strict reaction conditions. The primers and probes were checked for possible homologies to all sequences published in gene banks by sequence comparison analysis. The analytical specificity of **AmpliSens® MTC-FRT** PCR kit, which was found to be 100%, was checked by testing 67 reference strains and clinical isolates:

- 16 bacteria representative of the *mycobacterium tuberculosis complex* (*M.tuberculosis*, *M.bovis*, *M.bovis BCG*, etc.);
- 23 nontuberculous mycobacteria (*M.avium*, *M.fortuitum*, *M.gordonae*, *M.intracellulare*, *M.kansasii*, *M.marinum*, *M.paratuberculosis*, *M.phlei*, *M.scrofulaceum*, *M.xenopi*, *M.smegmatis*, *M.ulcerans*, *M.terrae*, etc.);
- Bacteria of other groups (*Brucella abortus*, *B.melitensis*, *B.ovis*, and *B.suis*; *Campylobacter jejuni*; *Chlamydia suis*; *Chlamydia abortus* and *Ch.felis*; *Cryptococcus neoformans*; *Enterobacter cloaca* and *E.faecalis*; *Enterococcus faecalis*; *Escherichia coli*; *Klebsiella pneumoniae*; *Listeria monocytogenes*; *Moraxella catarrhalis*; *Neisseria cinerea*, *N.elongata*, *N.flava*, *N.gonor*, *N.mucosa*, *N.sicca*, and *N.subflava*; *Pantoea agglomerans*; *Pasteurella tularensis*; *Proteus vulgaris* and *P.mirabilis*; *Pseudomonas aeruginosa*; *Salmonella enteritidis* and *S.typhi*; *Shigella flexneri* and *Sh.sonne*; *Staphylococcus aureus*; different clinical isolates of *S.aureus* MRSA, *S.faecalis*, *S.saprophyticus*; and different clinical isolates of *Streptococcus A*, *B*, *C*, *G*, *S.oralis*, and *S.pneumoniae*).

The analytical specificity of **AmpliSens® MTC-FRT** PCR kit was estimated by the absence of positive result of the non-tuberculosis bacterium DNA amplification and by the presence of positive result of the *Mycobacterium tuberculosis complex* DNA amplification.

⁴ Analysis can be performed without DNA extraction if washing fluids from environmental objects are added immediately to the reaction mixture for carrying out PCR analysis

14. REFERENCES

- Guidelines to **AmpliSens® MTC-FRT** PCR kit for qualitative detection of *Mycobacterium tuberculosis (MBT)* DNA – *Mycobacterium tuberculosis complex (MTC)* in biological material, cultures of microorganisms and environmental objects by the polymerase chain reaction (PCR) with real-time hybridization-fluorescence detection developed by Federal Budget Institute of Science “Central Research Institute for Epidemiology”.

15. QUALITY CONTROL

In compliance with Federal Budget Institute of Science “Central Research Institute for Epidemiology” ISO 13485-Certified Quality Management System, each lot of **AmpliSens® MTC-FRT** PCR kit has been tested against predetermined specifications to ensure consistent product quality.

List of Changes Made in the Instruction Manual

VER	Location of changes	Essence of changes
03.06.11 VV	Content Text	form AmpliSens® MTC-FRT PCR kit variant FRT, REF R-B57(RG,iQ,SC,Dt)-CE was deleted
	Catalogue number	REF R-B57(RG,iQ,SC,Dt)-CE was deleted
	Key to Symbols Used	The explanation of symbols was corrected
	Cover page, text	The name of Institute was changed to Federal Budget Institute of Science “Central Research Institute for Epidemiology”
21.06.11 VV	Cover page, text	The name of Institute was changed to Federal Budget Institute of Science “Central Research Institute for Epidemiology”
02.12.12 IvI	13.1. Sensitivity	The name of the strain, <i>M.tuberculosis</i> (H37 Ra strain), was added
17.04.13 ME	Content	Positive Control DNA <i>MTC</i> /STI (C+) was changed to Positive Control DNA <i>MTC</i> /STI (C+ _{MTCSTI})
14.06.13 PE	Cover page	IVD symbol was replaced with RUC symbol
	Key to symbols used	
28.04.15 PM	Text	Corrections according to the template. Grammar corrections
	1. Intended use	The phrase “The results of PCR analysis are taken into account in complex diagnostics of disease” was changed to “For research use only. Not for diagnostic procedures”
	8.1. DNA extraction	The clarifications were added
	9. Data analysis	The sections were rewritten
	10. Troubleshooting	
14. References	The reference to Guidelines was added	
27.08.18 TA	Content	The color of a reagent was specified
10.06.21 EM	Through the text	The text formatting was changed
	2. Principle of PCR detection	The table with targets was added
	Footer	The phrase “For research use only. Not for diagnostic procedures” was added
13.07.23 EM	3. Content Footer	REF R-B57(RG,iQ,SC,Dt)-CE was added

