

AmpliSens® BRCA1-FRT PCR kit



For Professional Use Only

Instruction Manual

KEY TO SYMBOLS USED

	Catalogue number		Caution
	Batch code		Contains sufficient for <n> tests
	Research Use Only		Use-by Date
	Version		Consult instructions for use
	Temperature limit		Keep away from sunlight
	Manufacturer		Positive controls of amplification
	Date of manufacture		Negative control of extraction
	Federal Budget Institute of Science "Central Research Institute for Epidemiology"		Negative control of amplification

1. INTENDED USE

AmpliSens® BRCA1-FRT PCR kit is an *in vitro* nucleic acid amplification test for determination of genetic susceptibility to cancer by qualitative detection of germline mutations (5382insC (rs80357906), 4153delA (rs80357711), 300T>G (C61G) (rs28897672) and 185delAG (rs80357914)) in the *BRCA1* gene in biological material (venous whole blood), using real-time hybridization-fluorescence detection of amplified products. The material for PCR is DNA samples extracted from test material.

Indications and contra-indications for use of the reagent kit

The reagent kit is used for determination of genetic susceptibility to cancer associated with mutations in the *BRCA1* gene, without distinction of presence of manifestation.

A negative test result does not rule out the presence of other mutations in the *BRCA1* gene and other genes, or the general population risk of cancer.

There are no contra-indications with the exception of cases when the material cannot be taken for medical reasons.

NOTE: For research use only. Not for diagnostic procedures

2. PRINCIPLE OF PCR DETECTION

Principle of testing is based on amplification of human genomic DNA fragments including 5382insC (rs80357906), 4153delA (rs80357711), 300T>G (C61G) (rs28897672) and 185delAG (rs80357914) mutations in the *BRCA1* gene with hybridization-fluorescence detection.

The *BRCA1* gene fragment with no mutation at position 5382 (5382 WT) is used as an endogenous internal control. The functional copy of the *BRCA1* gene must always be present in the test sample in a sufficient amount proportional to the number of human cells in the sample. Endogenous internal control allows not only to control all stages of the PCR analysis for each sample, but also to evaluate the adequacy of the material sampling and its storage.

Amplification of DNA fragments with the use of specific primers and Taq-polymerase enzyme are performed with the DNA samples obtained at the extraction stage. In the real-time PCR, the amplified product is detected with the use of fluorescent dyes. These dyes are linked to oligonucleotide probes, which bind specifically to the amplified product during thermocycling. The real-time monitoring of fluorescence intensities during the real-time PCR allows the detection of accumulating product without re-opening the reaction tubes after the PCR run.

AmpliSens® BRCA1-FRT PCR kit uses "hot-start", which greatly reduces the frequency of nonspecifically primed reactions. "Hot-start" is guaranteed by using chemically modified polymerase (TaqF). The chemically modified polymerase (TaqF) is activated by heating at 95 °C for 15 min.

The PCR kit contains the system for prevention of contamination by amplicons using the enzyme uracil-DNA-glycosylase (UDG) and deoxyuridine triphosphate (dUTP).

The results of amplification are registered in the following fluorescence channels:

Table 1

Channel for fluorophore	FAM	JOE	ROX	Cy5	Cy5.5
DNA-target	5382insC (rs80357906)	4153delA (rs80357711)	300T>G (rs28897672)	185delAG (rs80357914)	5382 WT (no mutation)
Target gene	<i>BRCA1</i>	<i>BRCA1</i>	<i>BRCA1</i>	<i>BRCA1</i>	<i>BRCA1</i>

3. CONTENT

AmpliSens® BRCA1-FRT PCR kit is produced in 1 form:

variant FRT-50 F, S-3901-1-CE

Variant FRT-50 F includes:

Reagent	Description	Volume, ml	Quantity
PCR-mix-FL BRCA1	clear liquid from colorless to blue grey colour	0.6	1 tube
PCR-buffer-C	colorless clear liquid	0.3	1 tube
Polymerase (TaqF)	colorless clear liquid	0.03	1 tube
C+ BRCA WT	colorless clear liquid	0.2	1 tube
C+ BRCA Het	colorless clear liquid	0.2	1 tube
TE-buffer	colorless clear liquid	0.2	1 tube
Negative Control (C-)*	colorless clear liquid	1.2	1 tube

* must be used in the extraction procedure as Negative Control of Extraction.

Variant FRT-50 F is intended for 55 reactions (including controls).

4. ADDITIONAL REQUIREMENTS

For sampling and pretreatment

- Reagent for pretreatment of whole peripheral and umbilical blood.
- Vacuum tubes for sampling, storage and transportation of blood samples.
- Sterile bilateral needles for vacuum tubes intended for venous blood collection.
- Disposable screwed or tightly closed polypropylene 1.5-ml tubes.
- Sterile pipette tips with aerosol filters (up to 100, 200 and 1000 µl).
- Tube racks.
- PCR box.
- Desktop centrifuge up to 12,000 g (suitable for Eppendorf tubes).
- Vortex mixer.
- Vacuum aspirator with flask for removing supernatant.
- Pipettes (adjustable).
- Refrigerator for 2–8 °C.
- Deep-freezer at the temperature from minus 24 to minus 16 °C.
- Reservoir to throw off and inactivate the material.
- Disposable powder-free gloves and a laboratory coat.

For DNA extraction and amplification

- DNA extraction kit.
- TE-buffer reagent, 5-ml tube.
- Sterile pipette tips with aerosol filters (up to 100, 200 and 1000 µl).
- Tube racks.
- Vortex mixer.
- PCR box.
- Real-time instruments (for example, Rotor-Gene Q (QIAGEN, Germany); CFX 96 (Bio-Rad, USA)).
- Disposable polypropylene tubes:
 - a) screwed or tightly closed 1.5-ml tubes for reaction mixture preparation.
 - b) thin-walled 0.2-ml PCR tubes with optical transparent domed or flat caps or strips of eight 0.2-ml tubes with optical transparent caps if a plate-type instrument is used;
 - c) thin-walled 0.2-ml PCR tubes with flat caps or strips of four 0.1-ml Rotor-Gene PCR tubes if a rotor-type instrument is used.
- Pipettes (adjustable).
- Refrigerator for 2–8 °C.
- Deep-freezer at the temperature from minus 24 to minus 16 °C.
- Reservoir for used tips.
- Disposable powder-free gloves and a laboratory coat.

5. GENERAL PRECAUTIONS

The user should always pay attention to the following:

- Use sterile pipette tips with aerosol filters and use a new tip for every procedure.
- Store all extracted positive material (specimens, controls and amplicons) away from all other reagents and add it to the reaction mix in a distantly separated facility.
- Thaw all components thoroughly at room temperature before starting an assay.
- When thawed, mix the components and centrifuge briefly.
- Use disposable protective gloves and laboratory cloths, and protect eyes while samples and reagents handling. Thoroughly wash hands afterwards.
- Do not eat, drink, smoke, apply cosmetics, or handle contact lenses in laboratory work areas.
- Do not use the PCR kit if the internal packaging was damaged or its appearance was changed.
- Do not use the PCR kit if the transportation and storage conditions according to the Instruction Manual were not observed.
- Do not use a kit after its expiration date.
- Dispose of all specimens and unused reagents in accordance with local regulations.
- Samples should be considered potentially infectious and handled in biological cabinet in compliance with appropriate biosafety practices.
- Clean and disinfect all samples or reagents spills using a disinfectant, such as 0.5 % sodium hypochlorite or another suitable disinfectant.
- Avoid inhalation of vapors, samples and reagents contact with the skin, eyes, and mucous membranes. Harmful if swallowed. If these solutions come into contact, rinse the injured area immediately with water and seek medical advice if necessary.
- While observing the conditions of transportation, operation and storage, there are no risks of explosion and ignition.
- Safety Data Sheets (SDS) are available on request.
- The PCR kit is intended for single use for PCR analysis of specified number of samples (see the section "Content").

- The PCR kit is ready for use in accordance with the Instruction Manual. Use the PCR kit strictly for intended purpose.
- Use of this product should be limited to personnel trained in DNA amplification techniques.
- Workflow in the laboratory must be one-directional, beginning in the Extraction Area and moving to the Amplification and Detection Area. Do not return samples, equipment and reagents in the area where the previous step was performed.



Some components of this kit contain sodium azide as a preservative. Do not use metal tubing for reagent transfer.

6. SAMPLING AND HANDLING

AmpliSens® BRCA1-FRT PCR kit is intended for analysis of the DNA extracted with DNA extraction kits from the biological material (venous whole blood).

Sampling

Venous whole blood should be taken with a disposable 0.8-1.1 mm diameter needle into the tube (special vacuum system) with EDTA as anticoagulant. After blood sampling the tube should be smoothly rotated several times for the thoroughly mixing with the anticoagulant. (Otherwise, blood will coagulate and DNA extraction will be impossible!) Place the tube in the rack after rotating.

Venous whole blood samples can be stored before pretreatment:

- at the temperature from 20 to 25 °C – for 6 hours after sampling of biological material;
- at the temperature from 2 to 8 °C – for 1 day.

Freezing of venous whole blood samples is unallowable!

Venous whole blood can be transported at 2–8 °C for 1 day.

Pretreatment

Venous whole blood is to be pretreated.

Transfer 1.0 ml of Hemolytic and 1.0 ml of venous whole blood to the 1.5-ml Eppendorf tube. Gently vortex the tube and leave it for 10 min, stirring occasionally. Centrifuge at 8,000 rpm for 2 min. Remove the supernatant using vacuum aspirator without touching the pellet. After washing the cell pellet should be white, only a small pinkish bloom on the pellet is allowed (the remains of the destroyed erythrocytes). The washing using Hemolytic may be repeated if necessary. The obtained leukocyte pellet must be immediately lysed (in case of extraction using **RIBO-prep** reagent kit add **300 µl of Solution for Lysis** and then extract DNA in accordance with the *Instruction Manual* enclosed to the **RIBO-prep** reagent kit **without adding Solution for Lysis once again**) or frozen.

The frozen pellet can be stored at the temperature from minus 24 to minus 16 °C for 2 weeks or at the temperature not more than minus 68 °C for 1 year.

In case of using other reagent kits for DNA extraction the test material should be pretreated if necessary in accordance with the *Instruction Manual* to the using reagent kit.

Interfering substances and limitations of using test material samples

The next samples are inapplicable for analysis:

- the venous whole blood samples, collected in the tubes with heparin as anticoagulant,
- the venous whole blood samples, containing blood clot or which has been exposed to freezing.

Potential interfering substances

Endogenous and exogenous substances that may be present in the biological material (venous whole blood) used for the study were selected to assess potential interference.

Venous whole blood samples with known genotypes without adding and with the addition of potential interfering substances were tested. Concentration of each potential interfering substance is specified in the Table 2. Each sample was tested in three repeats. Genotype of venous whole blood samples was confirmed by pyrosequencing with PyroMark series genetic analysis system.

Table 2

Type of potential interferent	Potential interferent	Tested concentration	Interference presence
Exogenous substances	Potassium EDTA	up to 2.0 mg/ml	Not detected
	Lithium heparin	from 12 IU/ml	Detected
Endogenous substances	Total bilirubin	210 µmol/l (upper limit of normal – 21 µmol/l)	Not detected
	Total cholesterol	78 mmol/l (upper limit of normal – 7.8 mmol/l)	Not detected
	Triglycerides	37.0 mmol/l (upper limit of normal – 3.7 mmol/l)	Not detected
	Hemoglobin	250 g/l (upper limit of normal – 170 g/l)	Not detected

7. WORKING CONDITIONS

AmpliSens® BRCA1-FRT PCR kit should be used at the temperature from 20 to 28 °C and relative humidity from 15 to 75 %.

8. PROTOCOL

8.1. DNA extraction

It is recommended to use the following nucleic acid extraction kits:

- RIBO-prep**.

If using the RIBO-prep reagent kit extract the DNA according to the manufacturer's protocol.

The volumes of reagents and samples when the DNA is extracted by RIBO-prep nucleic acid extraction kit:

The test sample is the pellet obtained after venous whole blood (**100 µl**) pretreatment using Hemolytic.

Add **100 µl of Negative Control (C–)** to the tube labeled C– (Negative Control of Extraction).

The volume of elution is **50 µl**.

8.2. Preparing PCR

8.2.1 Preparing tubes for PCR

The total reaction volume is **25 µl**, the volume of the **DNA** sample is **10 µl**.

The type of tubes depends on the PCR instrument used for analysis. Use disposable filter tips for adding reagents, DNA and control samples into tubes.

It is recommended to determine DNA concentration in obtained extracted samples and to dilute extracted DNA using TE-buffer to required concentration 10-30 ng per reaction (corresponds to concentration of extracted DNA 3x10⁵–9x10⁵ copies/m) (it is recommended to use TE-buffer reagent manufactured by FBIS CRIE).

1. Calculate the required quantity of each reagent for reaction mixture preparation. For one reaction:

- 10 µl of PCR-mix-FL BRCA1**,
- 5 µl of PCR-buffer-C**,
- 0.5 µl of polymerase (TaqF)**.

Prepare the reaction mixture for the total number of test and control samples plus some extra reactions. See numbers of control samples in item 7.

NOTE: Prepare the reaction mixture just before use.

- Thaw the tube with **PCR-mix-FL BRCA1**. Mix the content of the tubes with **PCR-mix-FL BRCA1**, **PCR-buffer-C** and **polymerase (TaqF)**. Sediment the drops by vortex.
- In a new tube prepare the reaction mixture. Mix the required quantities of **PCR-mix-FL BRCA1**, **PCR-buffer-C** and **polymerase (TaqF)**. Sediment the drops by vortex.

NOTE: The prepared mixture should be stored for no more than 2 hours.

4. Take the required number of the tubes or strips for PCR of DNA of test and control samples.

5. Transfer **15 µl** of the prepared reaction mixture to each tube. Discard the unused reaction mixture.

6. Transfer to the prepared tubes **10 µl of DNA samples** obtained by extraction from the test samples and diluted to the optimal concentration.

7. Carry out the control amplification reactions:

- C+ WT** – Add **10 µl of C+ BRCA WT** to the tube labeled C+ WT (Positive Control of Amplification)
- C+ Het** – Add **10 µl of C+ BRCA Het** to the tube labeled C+ Het (Positive Control of Amplification)
- NCA** – Add **10 µl of TE-buffer** to the tube labeled NCA (Negative Control of Amplification)
- C–** – Add **10 µl of the sample extracted from the Negative Control (C–) reagent** to the tube labeled C– (Negative control of Extraction).

8.2.2. Amplification

1. Create a temperature profile on your instrument as follows:

Table 3

AmpliSens unified amplification program for rotor-type ¹ instruments				
Step	Temperature, °C	Time	Fluorescent signal detection	Cycles
1	50	15 min	–	1
2	95	15 min	–	1
3	95	10 s	–	45
	60	20 s	FAM, JOE, ROX, Cy5, Cy5.5	

Any combination of the tests (including tests with reverse transcription and amplification) can be performed in one instrument simultaneously with the use of the unified amplification program for rotor-type instruments. If only the tests for DNA detection are performed in one instrument then the first step of reverse transcription (50 °C – 15 min) can be omitted for time saving.

NOTE:

Table 4

Amplification program for plate-type ² instruments				
Step	Temperature, °C	Time	Fluorescent signal detection	Cycles
1	95	15 min	–	1
2	95	10 s	–	45
	60	27 s	FAM, JOE, ROX, Cy5, Cy5.5	

Fluorescent signal is detected in the channels for the **FAM**, **JOE**, **ROX**, **Cy5** and **Cy5.5** fluorophores.

2. Adjust the fluorescence channel sensitivity according to the *Important Product Information Bulletin*.

3. Insert tubes into the reaction module of the device.

It is recommended to sediment drops from walls of tubes by short centrifugation before placing them into the instrument.

NOTE:

Insert empty tubes at the edges of reaction module in case of incomplete filling of plate-type instrument.

4. Run the amplification program with fluorescence detection.

5. Analyze results after the amplification program is completed.

9. DATA ANALYSIS

Analysis of results is performed by the software of the real-time PCR instrument used by measuring fluorescence signal accumulation in five channels:

Table 5

Channel for the fluorophore	FAM	JOE	ROX	Cy5	Cy5.5
Amplification product	5382insC (rs80357906)	4153delA (rs80357711)	300T>G (rs28897672)	185delAG (rs80357914)	5382 WT (no mutation)

Results are interpreted by the crossing (or not-crossing) the fluorescence curve with the threshold line set at the specific level that corresponds to the presence (or absence) of a *Ct* value of the DNA sample in the corresponding column of the results grid. The fluorescence curves of the sample should once cross the threshold line in the area of typical exponential growth of fluorescence. Characteristics of threshold line settings for plate-type and rotor-type instruments are specified in the *Important Product Information Bulletin*.

Principle of interpretation is the following:

Table 6

Results interpretation					Result
<i>Ct</i> value in the channel for the fluorophore					
FAM	JOE	ROX	Cy5	Cy5.5	
< boundary value	absent or > boundary value	absent or > boundary value	absent or > boundary value	< boundary value	5382insC (rs80357906) mutation is detected
absent or > boundary value	< boundary value	absent or > boundary value	absent or > boundary value	< boundary value	4153delA (rs80357711) mutation is detected
absent or > boundary value	absent or > boundary value	< boundary value	absent or > boundary value	< boundary value	300T>G (rs28897672) mutation is detected
absent or > boundary value	absent or > boundary value	absent or > boundary value	< boundary value	< boundary value	185delAG (rs80357914) mutation is detected
absent or > boundary value	absent or > boundary value	absent or > boundary value	absent or > boundary value	< boundary value	Detectable mutations in the BRCA1 gene are NOT detected
determined or absent	determined or absent	determined or absent	determined or absent	absent or > boundary value	Invalid* result

* In case of **invalid** result, the PCR analysis should be repeated for the corresponding test sample starting from the DNA extraction stage. If the same result was obtained once again, re-sampling of the material for analysis is recommended.

NOTE: Boundary *Ct* values are specified in the *Important Product Information Bulletin* enclosed to the PCR kit.

¹ For example, Rotor-Gene Q (QIAGEN, Germany).

² For example, CFX 96 (Bio-Rad, USA).

NOTE: It is recommended to confirm the result with additional analysis using sequencing methods for the samples with detected mutations. A negative test result does not rule out the presence of other mutations in the *BRCA1* gene and other genes, or the general population risk of cancer.

The result of the PCR analysis is considered reliable only if the results obtained for controls of extraction and amplification stages are correct (according to the Table 7 and the *Important Product Information Bulletin* enclosed to the PCR kit).

Table 7

Control	Stage for control	Results for controls				
		Ct value in the channel for fluorophore				
		FAM	JOE	ROX	Cy5	Cy5.5
C-	DNA extraction	absent or > boundary value	absent or > boundary value	absent or > boundary value	absent or > boundary value	absent or > boundary value
NCA	PCR	absent or > boundary value	absent or > boundary value	absent or > boundary value	absent or > boundary value	absent or > boundary value
C+ WT	PCR	absent or > boundary value	absent or > boundary value	absent or > boundary value	absent or > boundary value	< boundary value
C+ Het	PCR	< boundary value	< boundary value	< boundary value	< boundary value	< boundary value

10. TROUBLESHOOTING

Results of analysis are not taken into account in the following cases:

- The Ct value determined for the Positive Control of Amplification (C+ Het) is greater than the boundary value or absent in any of the specified channels for fluorophores (see Table 7). The amplification and detection should be repeated for all samples.
- The Ct value determined for the Positive Control of Amplification (C+ WT) is greater than the boundary value or absent in the channel for the Cy5.5 fluorophore. The amplification and detection should be repeated for all samples.
- If the Ct value determined for the Negative Control of Extraction (C-) and Negative Control of amplification (NCA) is less than the boundary value in the channels for the FAM and/or JOE, and/or ROX, and/or Cy5, and/or Cy5.5 fluorophores. The contamination of laboratory with amplification fragments or contamination of reagents, test samples is probable at any stage of PCR analysis. Measures for detecting and elimination of contamination source must be taken. The PCR analysis (beginning with the DNA extraction stage) should be repeated for all samples in which specific DNA was detected.
- If the Ct value determined for the test sample is not greater than the boundary value, whereas the area of typical exponential growth of fluorescence is absent (the graphic looks like approximate straight line). It is necessary to check the correctness of selected threshold line level. If the result has been obtained with the correct level of threshold line (base line), the amplification and detection should be repeated for this sample.
- If the Ct value determined for the test sample is not greater than the boundary value in two or more channels for the FAM, JOE, ROX and Cy5 fluorophores. The PCR analysis (beginning with the DNA extraction stage) should be repeated for all samples.
- If the Ct value determined for the test sample is less than the boundary value in the channel for the Cy5.5 fluorophore and the Ct value is less than the boundary value in one or more other detection channels (FAM and/or JOE, and/or ROX, and/or Cy5). The final fluorescence level in this/these channels is significantly lower than the final fluorescence level for C+ Het (see the *Important Product Information Bulletin* enclosed to the PCR kit). There may be no mutation detectable in this channel. The analysis (beginning with the amplification stage) should be repeated for the sample at the optimal concentration (see 8.2. *Preparing PCR*).

11. TRANSPORTATION

AmpliSens® BRCA1-FRT PCR kit should be transported at 2–8 °C for no longer than 5 days. PCR kit can be transported at 2–25 °C for no longer than 3 days.

12. STABILITY AND STORAGE

All components of the **AmpliSens® BRCA1-FRT** PCR kit are to be stored at 2–8 °C when not in use (except for PCR-mix-FL BRCA1, PCR-buffer-C and polymerase (TaqF)). All components of the **AmpliSens® BRCA1-FRT** PCR kit are stable until the expiry date stated on the label. The shelf life of reagents before and after the first use is the same, unless otherwise stated.

NOTE: PCR-mix-FL BRCA1, PCR-buffer-C and polymerase (TaqF) are to be stored at the temperature from minus 24 to minus 16 °C

NOTE: PCR-mix-FL BRCA1 is to be kept away from light

13. SPECIFICATIONS

13.1. Analytical sensitivity (limit of detection)

Table 8

Test material	The volume of sample for extraction, µl	Nucleic acid extraction kit	PCR kit	Analytical sensitivity (limit of detection), copies/ml
Venous whole blood	100	RIBO-prep	variant FRT-50 F	1x10 ⁴

The claimed features are achieved while respecting the rules specified in the section "Sampling and Handling".

13.2. Analytical specificity

The analytical specificity of **AmpliSens® BRCA1-FRT** PCR kit is ensured by the selection of specific primers and probes as well as stringent reaction conditions. The primers and probes have been checked for possible homologies to all sequences published in gene banks by sequence comparison analysis.

The PCR kit detects the presence of 5382insC (rs80357906), 4153delA (rs80357711), 300T>G (rs28897672) and 185delAG (rs80357914) mutations in the *BRCA1* gene. The analytical specificity was proved when testing 249 human genomic DNA samples with known genotypes for these mutations determined by sequencing with PyroMark Q24 pyrosequencer.

The nonspecific responses and incorrect mutation detection were not observed while testing the samples.

The information about interfering substances is specified in the *Interfering substances and limitations of using test material samples*.

13.3. Repeatability and reproducibility

Repeatability and reproducibility were determined by testing of positive and negative model samples. Quality control samples (QCS) were used for testing in concentration of 1x10⁴ copies/ml containing mutations in the *BRCA1* gene as positive sample and not containing mutations in the *BRCA1* gene (wild type) as negative sample. Each sample was tested in 10 repeats.

Repeatability conditions included testing in the same laboratory, by the same operator, using the same equipment within a short period of time. Reproducibility conditions included testing different lots of reagent kit in two independent laboratories, by different operators, using different equipment. The results are presented in Table 9.

Table 9

Mutation	Sample type	Repeatability		Reproducibility	
		Number of samples	Agreement of results, %	Number of samples	Agreement of results, %
5382insC	Positive	10	100	40	100
	Negative	10	100	40	100
4153delA	Positive	10	100	40	100
	Negative	10	100	40	100
300T>G	Positive	10	100	40	100
	Negative	10	100	40	100
185delAG	Positive	10	100	40	100
	Negative	10	100	40	100

14. REFERENCES

- NCCN Clinical Practice Guidelines in Oncology Genetic/Familial High-Risk Assessment: Breast, Ovarian, and Pancreatic Version 2.2021 // National Comprehensive Cancer Network, Inc. 2020.
- Karami F., Mehdipour P. A comprehensive focus on global spectrum of BRCA1 and BRCA2 mutations in breast cancer. // Biomed Res. 2013. doi: 10.1155/2013/928562.

15. QUALITY CONTROL

In compliance with Federal Budget Institute of Science "Central Research Institute for Epidemiology" ISO 13485-Certified Quality Management System, each lot of the **AmpliSens® BRCA1-FRT** PCR kit has been tested against predetermined specifications to ensure consistent product quality.

AmpliSens®



Federal Budget Institute of Science "Central Research Institute for Epidemiology"
3A Novogireevskaya Street
Moscow 111123 Russia